

**CERTIFICATE OF COMPLIANCE
RoHS Declaration**

MANUFACTURER: Illumina Inc.,
5200 Illumina Way
San Diego, CA 92122 USA

Reference Number: 20005715
Product Name: NextSeq™ 550Dx Instrument
Model Number(s): 20005715

Commission Delegated Directive (EU) 2015/863 of 31 March 2015 amending Annex II to Directive 2011/65/EU of the European Parliament and of the Council as regards the list of restriction substances.

This document declares that Illumina Inc. *In Vitro* Diagnostic (IVDs)/medical devices are in conformity with Directives: 2015/863 and 2011/65 of the European Parliament and of the Council on the restriction of the use of certain hazardous substances in electrical and electronic equipment with exemptions.

This declaration is based on analysis of raw materials used in the manufacturing process and suppliers' declarations.

This declaration of conformity is issued under the sole responsibility of the manufacturer.

Annex II Restricted substances referred to in Article 4(1) and maximum concentration values tolerated by weight in homogeneous materials are:

Restricted Hazardous Substances	
Lead (Pb) – 0.1%	Polybrominated diphenylethers (PBDE) – 0.1%
Mercury (Hg) – 0.1%	Bis(2-Ethylhexyl) phthalate (DEHP) – 0.1%
Cadmium (Cd) – 0.01%	Benzyl butyl phthalate (BBP) – 0.1%
Hexavalent chromium (Cr VI) – 0.1%	Dibutyl phthalate (DBP) – 0.1%
Polybrominated biphenyls (PBB) – 0.1%	Diisobutyl phthalate (DIBP) – 0.1%

Name/Title	Signature	Date
E. Joseph McMullen Sr. Director, Regulatory Affairs Department		10 Mar 2021