



Declaration of Conformity

Illumina, Inc. hereby declares under its sole responsibility that the product(s) listed are in conformity to the EMC Directive [2014/30/EU], Low Voltage Directive [2014/35/EU], RoHS Directive [2011/65/EU] as amended by Commission Delegated Directive (EU) 2015/863.

MANUFACTURER: Illumina, Inc
ADDRESS: 5200 Illumina Way
San Diego, CA 92122, USA

FACTORY LOCATION:
Illumina Singapore Pte. Ltd
North Tech Lobby 3 #02-13118
29 Woodlands Industrial Park E1
Singapore, 757716

PRODUCT TYPE: Microarray Scanner
MODEL: iScan
CE MARK AFFIXED: 2008

AUTHORIZED EU REPRESENTATIVE:
Illumina Netherlands B.V.
Steenoven 19
5626 DK Eindhoven
The Netherlands

The construction of the product is in compliance with the following harmonized and/or consensus standards.

EN 61010-1:2010/A1 (3 rd Edition)	<i>Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements</i>
EN 61010-2-081:2015	<i>Particular requirements for automatic and semi-automatic laboratory equipment for analysis and other purposes</i>
EN 60825-1:2014	<i>Safety of laser products - Part 1: Equipment classification and requirements</i>
EN 61326-1:2020 (Class A)	<i>Electrical equipment for the measurement, control and Laboratory use – EMC Requirements Part1, Class A</i>
EN 55011:2016+ A11	<i>Industrial, scientific and medical equipment - Radio-frequency disturbance characteristics - Limits and methods of measurement</i>
EN 61000-3-2:2019 + AMD1:2020	<i>Electromagnetic compatibility (EMC) - Part 3-2: Limits - Limits for harmonic current emissions (equipment input current ≤16 A per phase)</i>
EN 61000-3-3:2013/A2:2021	<i>Electromagnetic compatibility (EMC) - Part 3-3: Limits - Limitation of voltage changes, voltage fluctuations and flicker in public low-voltage supply systems, for equipment with rated current ≤16 A per phase and not subject to conditional connection</i>
EN 61000-4-8:2010	<i>Electromagnetic compatibility (EMC) - Part 4-8: Testing and measurement techniques - Power frequency magnetic field immunity test</i>
EN 63000:2018	<i>Technical documentation for the assessment of electrical and electronic products with respect to the restriction of hazardous substances</i>

Illumina declares the product listed above is in compliance with RoHS Directive 2011/65/EU, as amended by (EU) 2015/863 on the restriction of the use of certain hazardous substances in electrical and electronic equipment.

This declaration is based on analysis of raw materials used in the manufacturing process and supplier's declarations.

Lead (0,1%)	Polybrominated diphenylethers (PBDE) (0,1%)
Mercury (0,1%)	Bis(2-Ethylhexyl) phthalate (DEHP) (0,1%)
Cadmium (0,01%)	Benzyl butyl phthalate (BBP) (0,1R%)
Hexavalent chromium (0,1%)	Dibutyl phthalate (DBP) (0,1%)
Polybrominated biphenyls (PBB) (0,1%)	Diisobutyl phthalate (DIBP) (0,1%)

Annex III exemptions are applied.

Authorized by:

Karen Gutekunst Electronically signed by: Karen Gutekunst
Reason: Approver
Date: Dec 6, 2023 11:37 PST

06-Dec-2023

Karen Gutekunst
VP, Regulatory Affairs

Date

Revision: 00